## **VERIFICATION OF CALIBRATION REPORT**

of DataMaster cdm Breath Test Instrument State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DEC 29 2009

	DataMaster cdm S/N130195
Super	visor/Operator Performing the Verification Procedure:
A	Name Robert J. Ely ID# 2681 Date 12-19-2009
	Agency Craig Police Department Phone # (907) 826-3330
	Instrument Location506 2 <sup>nd</sup> Street, Craig Alaska 99921
	Alco S/N 77286 Target Value079 High Pressure 1000 PSI
	Alco Test Value Average
	Signature fourt 1. (OVER)
	(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this

become and sworn before me this I

(Notary Seal Stamp

Carolyn M. Noland

Notary Public, State of Alaska Commission Expires with Office

## VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

<u>13</u>0195 DataMaster cdm S/N

(CONTINUED FROM FRONT PAGE)

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY DATAMASTER cdm 130195 ~

DECEMBER 19, 2889

C

OPERATOR'S HOME: ELY, RJ OPERATOR'S NUMBER: 2681 SUBJECT'S LAST HAVE: VERIFICATION OF CAL SUBJECT'S FIRST NAME/MI : VER OF CAL O.L. #: 123456789 DEPT/AGENCY: CRG1 CASE/REPORT: 2009-XXXX TEST TYPE: V ALCO TARGET VALUE: .879 ALCO S/N: 77286

- BREATH ANALYSIS -

,079 POJUSTED FOR 39.53 in 13:33 ALCO TARGET 13:34 BLANK TEST 13:34 INTERNAL STANDARD PLCO TV 38.53 in VERIFIED .885 .000 BLANK TEST SUBJECT SAMPLE .000 BLANK TEST .085 ALCO TV 38.53 in 13:38 .000 BLANK TEST

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY DATAMPSTER cda 130195

DECEMBER 19, 2009 TIME 13:32

- DIAGNOSTIC CHECK --

OKRY

COMPUTER:

OKAY

SOFTWARE DATE: 02/20/01

HEATERS

PROGRAM:

SAMPLE CHAMBER: 49c 41c BREATH TUBE:

39.53 in BAROMETER:

OKAY FLOW DETECTOR:

OKRY HIGH SPEED:

OKRY DETECTOR:

OKAY FILTERS:

NKRY QUARTZ STANDARD:

OKAY CALIBRATION:

PRINTER TEST !"#\$2&?'()\*+,-,'0123456789:;(=)?@GBCDEF6 HIJKL#NOPORSTUWAXYZ[\]^\_'abcdef9hijklamo porstuvexyz(!)~